

AMS
TLW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010520 Worksheet Start Date 4/2/2014
Location Albemarle County Regional Jail
Address 160 Peregrory Ln Charlottesville VA 22902
DFS Technician Heather Stanton License No. 18910

☐ Laboratory

☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 747 Reference Barometer (mm HG) 749
Reference Barometer(RB)Serial # 009113 RB Calibration Due 5/8/2014

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.)

AG320502-22

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☐ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010520

Certification Date

☐ Calibrated☐ Certified☐ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

4/2/14

Issuing Analyst


msk

Date

4/3/14

Ans
TLW



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED	NAME OF COURT
INSTRUMENT, TEST,	DFS

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY		AGENCY
STANTON, HEATHER, M		DFS Central Lab
DFS LICENSE NUMBER	LICENSE EXPIRES	DATE TEST CONDUCTED
18910	10/01/2014	04/02/2014
TEST EQUIPMENT NUMBER		
010520		

RESULTS: TIME SAMPLE TAKEN 12:32 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 02 Apr 2014 at 15:07

HLN
TLN

Test Results

Instrument Serial Number 010520

Test # 001424

Subject Test

Test Location 1 Department of
Test Date 02 Apr 2014

Test Location 2 Forensic Science
Test Time 12:26
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 02 Apr 2014 End Time 12:34

Result Time 12:32

Result Date 02 Apr 2014

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:26

Data Type BLK

Sample Value 0.000

Sample Time 12:27

Data Type CHK

Sample Value 0.100

Sample Time 12:28

Data Type BLK

Sample Value 0.000

Sample Time 12:29

Data Type SUBJ

Sample Value 0.000

Sample Time 12:30

Data Type BLK

Sample Value 0.000

Sample Time 12:31

Data Type SUBJ

Sample Value 0.000

Sample Time 12:32

Data Type BLK

Sample Value 0.000

Sample Time 12:33

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG320502-22

Standard Expiration Date 07/24/2015

Tank Pressure 943

Barometric Pressure 747 mmHg

Blow Sample Number 1 Blow Duration 3.39 sec

Blow Volume 1755 cc

End-of-Blow Time 12:30

Blow Sample Number 2 Blow Duration 3.44 sec

Blow Volume 1780 cc

End-of-Blow Time 12:32

Tamper Evident Stamp 57a0695e

Test Status Code 0

Test Status Success